

BULLETIN #1 of 2026. April 29, 2026

IBSOPHONE



DISCONTINUANCE OF PRODUCT

Art. No: IBS0079-3 (30709-01) Bone Conductor Pouch

Products Affected: IBS0079-3 (30709-01) Bone Conductor Pouch

The Bone Conductor Pouch does not always place the Bone Conductor at an optimal position, and the neoprene tends to dampen the vibrations, which sometimes reduce the quality of the sound.

The discontinuance is not based on any "safety" factors whatsoever.

Effective now the products below are discontinued and not available for purchase.

IBS0079-3 (30709-01) Bone Conductor Pouch

Replacement product: IBS00175-01 (57040-51) Bone Conductor Holder, Silicone (Yellow)

Please contact you authorized Ibsophone dealer or e-mail sales@ibsophone.se for more information.